

資料提供招請に関する公表

次のとおり物品の購入を予定していますので、当該導入に関して資料等の提供を招請します。

令和 7 年 1 月 24 日

独立行政法人国立病院機構

四国こどもとおとなの医療センター院長
前田 和寿

◎ 調達機関番号 597 ◎ 所在地番号 37

○ 第『施設入力』号 (No. 『施設入力』)

1 調達内容

(1) 品目分類番号 22、31

(2) 導入計画物品及び数量

マルチスライスCT装置 一式

(3) 調達方法 購入等

(4) 導入予定時期 令和 7 年度第 4 ・ 四半期

(5) 調達に必要なとされる基本的な要求要件

① 1 スキャン 320 列以上で撮影することができること。

② 1 回の撮影の範囲が 160mm 以上できること。

③ 頭部及び心臓を 1 回転で撮影できること。

2 資料及びコメントの提供方法 上記 1 (2) の

物品に関する一般的な参考資料及び同 (5) の要

求要件等に関するコメント並びに提供可能なライブラリーに関する資料等の提供を招請する。

(1) 資料等の提供期限 令和7年2月25日17時00分（郵送の場合は必着のこと。）

(2) 提出先 〒765-8507 香川県善通寺市仙遊町2-1-1 独立行政法人国立病院機構四国こどもとおとなの医療センター企画課長玉井健一 電話 0877-62-1000

3 説明書の交付 本公表に基づき応募する供給者に対して導入説明書を交付する。

(1) 交付期間 令和7年1月24日から令和7年2月25日まで。

(2) 交付場所 上記2(2)に同じ。

4 その他

(1) 本公表は検討段階であり、変更することがあり得る。

(2) この情報提供依頼の詳細は情報提供依頼書による。

6 Summary

(1) Classification of the products to be procured: 22, 31

(2) Nature and quantity of the products to be purchased : Multi-slice Computed Tomography System 1set

- (3) Type of the procurement : Purchase
- (4) Basic requirements of the procurement
: ① One scan can be photographed in more than 320 rows
② The range of a single shot must be 160 mm or more
③ The head and heart must be photographed in one rotation
- (5) Time-limit for the submission of the requested material : 5:00 P.M February 25, 2025
- (6) Contact point for the notice : Kenichi Tamai, Director, Planning Division, National Hospital Organization Shikoku Medical Center for Children and Adults, 2-1-1, Senyu-cyou, Zentsuujishi, Kagawa-ken, 765-8507, Japan. TEL 0877-62-1000